

AD-A068 955

NAVY UNDERWATER SOUND REFERENCE LAB ORLANDO FLA
ELECTROACOUSTIC MEASUREMENTS ON MASSA TR-208A TRANSDUCERS SERIA--ETC(U)
JUN 70


F/G 17/1

UNCLASSIFIED

USRD-CALIBRATION-3123

NL

| OF |
AD
A068955



END
DATE
FILMED
6-79
DDC

MOST Project 4

14) USRD-Calibration 3123
K03-30.101/3883

LEVEL II

DDVI LIBRARY COPY

1

nw

AD A068955

DDC FILE COPY

6
Electroacoustic Measurements on
Massa TR-208A Transducers
Serials 24102 and 24141

11
18 Jun 1976

12 9p.

DDC
RECEIVED
MAY 24 1979

D

NAVAL RESEARCH LABORATORY
Underwater Sound Reference Division

P. O. Box 8337

Orlando, Florida 32806

DISTRIBUTION STATEMENT A

Approved for public release;
Distribution Unlimited

254 300 mt



000815

LEVEL II

Naval Research Laboratory
UNDERWATER SOUND REFERENCE DIVISION
P.O. Box 8337, Orlando, Florida 32806

VPB/HD/hs
K03-30.101/3883
18 June 1970

CALIBRATION REPORT No. 3123

Subj: Massa TR-208A transducers serials 24102 and 24141; electroacoustic measurements on

Encl: (1) Drawings USRD 62724 through 62727 and 45001
(2) Tables 1 and 2

1. Electroacoustic measurements were provided on the subject transducers as the result of a telephone conversation on 17 April 1970 between Mr. James Stockton of the Applied Research Laboratory, University of Texas, and Mr. Ivor Groves of USRD.

2. Transmitting current response, equivalent series impedance, and high-power electroacoustic characteristics were measured in the Anechoic Tank Facility during the period June 1969 to June 1970. The measurement conditions and the results are shown in enclosures (1) and (2). given.

3. Orientation was according to the method described for a piston on drawing USRD 45001; the serial number was in the direction of the +Z axis.

4. All measurements reported here that are described in "American Standard Procedures for Calibration of Electroacoustic Transducers, Particularly Those for Use in Water, Z24.24-1957" were made in accordance therewith.

V. P. Benedetti
V. P. BENEDETTI

Copy to:
NAVSHIPSYSKOM (Code 00V1D)(1)
ARI, U. Texas (James Stockton)(1)
USRD (Code 8270)(1)
(Code 8280)(1)
NRL Wash (Code 2020)(1)
(Code 1065)(1)

CLASSIFIED BY	
UFC	DATE CLASSIFIED
UFC	BY
CLASSIFICATION	
Per Hx. on file	
BY	
DISTRIBUTION/AVAILABILITY CODE	
Dist.	AVAIL. CODE/US
A	

DISTRIBUTION STATEMENT A
Approved for public release;
Distribution Unlimited

DDC
RECEIVED
MAY 24 1979
D

3
Naval Research Laboratory
UNDERWATER SOUND REFERENCE DIVISION
P. O. Box 8337, Orlando, Florida 32806

USRD No. 62724

~~3883~~ 3883

Date: - Jun 1969

TRANSMITTING CURRENT RESPONSE

Massa TR-208A Transducer Serial 24102


Pressure at one meter per ampere measured at end of 30-ft cable

Unbalanced

0 and 25 psig

Water temp: 3 °C

MEASUREMENTS MADE IN AC-
CORDANCE WITH AMERICAN
STANDARD Z.24.24-1957



2
Naval Research Laboratory
UNDERWATER SOUND REFERENCE DIVISION
P. O. Box 8337, Orlando, Florida 32806

USRD No. 62725

~~3883~~ 3883

Date: Apr 1970

TRANSMITTING CURRENT RESPONSE

Massa TR-208A Transducer Serial 24102


Pressure at one meter per ampere measured at end of 30-ft cable

Unbalanced

25 psig

Water temp: 25 °C

MEASUREMENTS MADE IN AC-
CORDANCE WITH AMERICAN
STANDARD Z 24.24-1957



Naval Research Laboratory
UNDERWATER SOUND REFERENCE DIVISION
P. O. Box 8337, Orlando, Florida 32806

USRD No. 62726

~~3883~~ 3883

Date: Jun 1969

TRANSMITTING CURRENT RESPONSE

Massa TR-208A Transducer Serial 24141


Pressure at one meter per ampere measured at end of 30-ft cable

Unbalanced

0 and 25 psig

Water temp: 3 °C

MEASUREMENTS MADE IN AC-
CORDANCE WITH AMERICAN
STANDARD Z.24.24-1957



Naval Research Laboratory
UNDERWATER SOUND REFERENCE DIVISION
P. O. Box 8337, Orlando, Florida 32806

USRD No. 62727

~~3883~~ 3883

Date: Apr 1970

TRANSMITTING CURRENT RESPONSE

Massa TR-208A Transducer Serial 24141

Pressure at one meter per ampere measured at end of 30-ft cable

Unbalanced

25 psig

Water temp: 25°C

MEASUREMENTS MADE IN AC-
CORDANCE WITH AMERICAN
STANDARD Z 34.24-1957




Table 1

EQUIVALENT SERIES IMPEDANCE

Massa TR-208A Transducers
Unbalanced; 30-ft cable
Hydrostatic pressure: 25 psig

MEASUREMENTS MADE IN ACCORDANCE WITH
FEDERAL BUREAU OF STANDARDS
271-241-57

Water temp (°C)	Freq (kHz)	Serial 24102		Serial 24141	
		R (ohms)	X (ohms)	R (ohms)	X (ohms)
3	3.5	20.7	-46.8	21.8	-52.4
	4.0	22.0	-18.3	22.7	-22.0
	4.5	30.9	-2.73	31.5	-5.70
	5.0	49.6	1.86	48.8	-2.20
	5.5	70.7	-19.6	67.1	-22.3
	6.0	55.1	-51.9	52.7	-50.4
	6.5	28.8	-51.4	28.4	-50.3
	7.0	17.0	-42.0	17.0	-41.5
25	3.5	20.7	-51.9	20.5	-56.2
	4.0	20.0	-22.8	20.1	-24.6
	4.5	26.5	-6.72	28.0	-6.90
	5.0	42.8	5.64	49.0	3.23
	5.5	79.6	-10.6	75.5	-26.5
	6.0	54.2	-56.9	46.6	-54.0
	6.5	25.7	-51.4	24.3	-48.5
	7.0	15.2	-41.3	15.7	-40.1

Table 2

Measurements made in accord-
ance with American Standard
Z39.1-61

ELECTROACOUSTIC CHARACTERISTICS

Massa TR-208A Transducers

Unbalanced; 30-ft cable
Hydrostatic pressure: 25 psig
Water temp: 3°C
High-power data


Duty cycle: 6% (3-msec pulse at 20 pps)

Serial	Freq (kHz)	Voltage (rms V)	Current (rms A)	Phase angle (deg)	Driving power (W)	Sound pressure level (dB re 1 μ bar at 1 m)	R (ohms)	X (ohms)
24102	5.0	75.0	1.51	2	113	94.7	49.6	1.73
		150	2.98	2	447	100.7	50.3	1.76
		157	3.17	2	497	101.3	49.5	1.73
	5.5	78.0	1.08	-16	81.0	93.8	69.4	-19.9
		158	2.19	-17	331	99.9	69.0	-21.1
		190	2.57	-16	469	101.4	71.1	-20.4
		194	2.68	-16	500	101.7	69.6	-20.0
	6.0	77.0	1.07	-43	60.3	92.9	52.6	-49.1
		156	2.20	-43	251	99.1	51.9	-48.4
		208	2.94	-43	447	101.6	51.7	-48.3
		220	3.09	-43	497	102.1	52.1	-48.6
24141	5.0	73.0	1.52	-5	111	94.7	47.8	-4.18
		105	2.13	-4	223	97.8	49.2	-3.44
		147	3.02	-3	443	100.7	48.6	-2.55
		156	3.20	-3	499	101.3	48.7	-2.55
	5.5	75.0	1.07	-18	76.3	93.2	66.7	-21.6
		105	1.52	-19	151	96.3	65.3	-22.5
		148	2.12	-19	297	99.3	66.0	-22.7
		187	2.66	-18	473	101.3	66.9	-21.7
		194	2.77	-18	511	101.6	66.6	-21.6
	6.0	77.0	1.07	-42	61.2	92.8	53.5	-48.2
		109	1.55	-42	126	95.9	52.3	-47.1
		153	2.18	-43	244	98.8	51.3	-47.9
		214	3.09	-43	484	101.9	50.7	-47.2
		220	3.18	-43	512	102.1	50.6	-47.2

1 July 1966

COORDINATE SYSTEM FOR TRANSDUCER ORIENTATION

The left-handed coordinate system of the American Standard Procedures for Calibration of Electroacoustic Transducers Particularly Those for Use in Water, Z24.24-1957, is used. The transducer is fixed with respect to the coordinate system and has its acoustic center at the origin. The angle ϕ is equivalent to the azimuth angle in sonar operation.



PLACEMENT OF TRANSDUCER IN COORDINATE SYSTEM

Transducer Type	Transducer Orientation in Coordinate System
Point, or Spherical	Points on surface that coincide with the X and Z axes shall be specified.
Cylindrical, or Line	The axis of the cylinder or line shall coincide with the Z axis. A reference mark in the XZ plane and in the direction of the positive X axis will be specified.
Plane, or Piston	The plane or piston face shall be in the YZ plane with the X axis normal to the face at its acoustic center. A reference mark in the XZ plane and in the direction of the positive Z axis will be specified.
Other Configurations	Orientation shall be shown by sketch or description. This category includes line and piston types of transducers operated in an orientation other than those specified above.

ORIENTATIONS FOR RESPONSE AND DIRECTIVITY MEASUREMENTS

Response. The calibration measurements are made for sound propagated parallel to the positive X axis ($\phi = 0$, $\theta = 90$), unless otherwise specified on the response curve.

Directivity. The plane of the pattern is specified, and the following conventions are observed, if another orientation is not specified on the pattern:

XY Planes: The positive X axis ($\phi = 0$, $\theta = 90$) coincides with the zero-degree direction on the pattern and the positive Y axis ($\phi = 90$, $\theta = 90$) is at 90 degrees measured in a clockwise direction. Rotation is around the Z axis; the positive Z axis is directed upward from the plane of the paper.

XZ Planes: The positive X axis coincides with the zero-degree direction and the positive Z axis ($\theta = 0$) is at 90 degrees measured in a clockwise direction. Rotation is around the Y axis; the negative Y axis is directed upward from the plane of the paper.

YZ Planes: The positive Y axis coincides with the zero-degree direction and the positive Z axis is at 90 degrees measured in a clockwise direction. Rotation is around the X axis; the positive X axis is directed upward from the plane of the paper.